



Leibniz Institute for **Tropospheric Research**

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

CPC Model:	TSI CPC 3750
CPC Serial Number:	3750193601
Customer:	TSI Instruments Ltd.
Description:	Calibration of a Condensation Particle Counter (CPC, Model 3750)
Date of Calibration:	November 13, 2019

Summary of Intercomparison:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 97% efficiency at 40 nm. The Dp50 is at 7.1 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Certificate / Reference: WCCAP

Date of issue: November 13, 2019 Signature:

Reviewed by: TROPOS

Name: Kay Weinhold

Page 1/4

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de





World Calibration Centre for Aerosol Physics



Leibniz Institute for Tropospheric Research

Date of arrival of instrument in calibration lab: Instrument: Model and serial number of instrument:

Result of physical inspection: Result of functional test:

Internal parameters of instrument

Model and identification number of aerosol electrometer:

Electrometer calibration certificate:

Corrections of electrometer, for instance, differing flow rate:

Software for recording:

Date of calibration: Lab temperature and pressure: Measured aerosol flow rate of CPC: Uncertainty in measured flow rate: Flowmeter used:

Particles and gases used for calibration: Method of particle generation: Zero measurement of instrument: November 11, 2019 Condensation Particle Counter CPC 3750 S/N 3750193601

no damages functional test successful, no problems

nominal flow rate 1.0 l/min

TSI Electrometer Model 3068, S/N 70838596

September 5, 2018, calibrated at PTB Braunschweig

Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.000 l/min LabView 2010; National Instruments; Program "LabCount.vi"

November 13, 2019 24.5°C, 982.5 mbar 0.988 l/min 3% Gilian Gilibrator V; S/N 1711008-S, January, 2018 silver particles and nitrogen tube furnace generator 0 particles/cm³ in 5 minutes

_ Results (using pulse output).						
Particle size (nm)	40	30	20	15	10	
Number concentration (cm-3)	1258	1243	1465	1214	1165	
Counting efficiency η	0.97	0.97	0.97	0.95	0.81	
Particle size (nm)	09	08	07	06		
Number concentration (cm-3)	1214	745	655	362		
Counting efficiency η	0.73	0.61	0.46	0.26		

Results (using pulse output):

Page 2/4

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de






Fig. 1: Counting efficiency for CPC 3750 S/N 3750193601 against aerosol electrometer 3068 S/N 70838596; silver particles between 6 and 40 nm were used for calibration; the calculated Dp50 is 7.10 nm.

Status information:

Status	T SAT	T CON	Τ ΟΡΤ	T CAB	P AMB	P VAC	
from display	39.0	18.0	40.0	24.3	99.0	75.4	
Status	P OR	P NO	Laser	LV	flow	P INLET	
from display	72.4	2.44	41	full	0.988	-0.4	

Results (using pulse output):

without coincidence correction						
Concentration EM in #/cm ³	66849	58164	50013	43197	32294	
Number concentration without						
coincidence correction (cm-3)	51580	46149	40692	35813	27615	
Counting efficiency η	0.77	0.79	0.81	0.83	0.85	
Concentration EM in #/cm ³	19887	11244	5127	1177		
Number concentration without						
coincidence correction (cm-3)	17688	10327	4853	1132		
Counting efficiency η	0.89	0.92	0.94	0.96		

Page 3 / 4

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de







Leibniz Institute for Tropospheric Research

World Calibration Centre for Aerosol Physics

with coincidence correction							
Concentration EM in #/cm ³	66849	58164	50013	43197	32294		
Number concentration with							
coincidence correction (cm-3)	66482	58023	49860	42923	31929		
Counting efficiency η	1.00	1.00	0.99	0.99	0.99		
Concentration EM in #/cm ³	19887	11244	5127	1177			
Number concentration with							
coincidence correction (cm-3)	19586	11167	5181	1231			
Counting efficiency η	0.98	0.99	1.01	1.04			




Fig. 2: Linearity test for TSI CPC 3750 SN 3750193601 against aerosol electrometer 3068 SN 70838596; silver particles with a diameter of 30 nm were used for number concentrations between 1000 and 70000 particles per cm³.

Date of issue: November 13, 2019

Reference: TSI electrometer, model 3068, SN 70838596 Reviewed: TROPOS / Kay Weinhold

Page 4 / 4

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de

