



European Centre for Aerosol Calibration
Administrating Office
Leibniz Institute for Tropospheric Research
Permoserstr. 15
04318 Leipzig

Geiger, Jutta
NRW State Agency for Nature, Environment and Consumer Protection
Leibnitzstrasse 10, D-45659 Recklinghausen
Germany

TNA – ECAC Confirmation of Intercomparison/Calibration

We confirm that in the context of ACTRIS Transnational Access the following inter-comparison was performed at the **European Center for Aerosol Calibration (ECAC)**:

Project number: **OCEC-2021-1**

Inter-comparison/Calibration date: **March 2021**

- ☒ Instrument: **Sunset Lab OCEC analyser** (serial number Model 4L, Serial Nr. 337-2013)
☐ Laboratory: *(institute)*

The Inter-comparison/Calibration was:

- ☒ **Successful**
☐ Unsuccessful
☐ Another calibration is recommended (see calibration report).
The PI has to submit a new TNA application.

Statistics for data submission to EBAS

Statistics	QA_bias %	QA_variability %	Systematic
TC	-1.8%	3.2%	no
OC	-0.9%	2.8%	no
EC	-8.7%	30.2%	no

Number of RWD (Researcher Working Days) : 3

Ispira, 09.06.2021

p.p. J.P. Putaud

Signature

Head of ERLAP, or ACMCC