



European Centre for Aerosol Calibration
Administrating Office
Leibniz Institute for Tropospheric Research
Permoserstr. 15
04318 Leipzig

Godec, Ranka
Institute for Medical Research and Occupational Health
Ksaverska cesta 2, POB 291, HR-10001 Zagreb,
CROATIA

TNA – ECAC Confirmation of Intercomparison/Calibration

We confirm that in the context of ACTRIS Transnational Access the following inter-comparison was performed at the **European Center for Aerosol Calibration (ECAC)**:

Project number: **OCEC-2021-1**

Inter-comparison/Calibration date: **March 2021**

- Instrument: **Sunset Lab OCEC analyser** (serial number 5L 466-235)
 Laboratory: (*institute*)

The Inter-comparison/Calibration was:

- Successful**
 Unsuccessful
 Another calibration is recommended (see calibration report).
The PI has to submit a new TNA application.

Statistics for data submission to EBAS

Statistics	QA_bias %	QA_variability %	Systematic
TC	-1.2	1.7	no
OC	1.9	2.4	no
EC	-20.0	17.0	low

Number of RWD (Researcher Working Days) : 3

Ispra, 09.06.2021
p.p. J.P. Putaud
Signature
Head of ERLAP, or ACMCC