



European Centre for Aerosol Calibration
Administrating Office
Leibniz Institute for Tropospheric Research
Permoserstr. 15
04318 Leipzig

Veronique PONT
Laboratoire d'Aérodologie UMR 5560
Observatoire Midi Pyrénées, 14 avenue Edouard Belin, 31400 Toulouse,
France

TNA – ECAC Confirmation of Intercomparison/Calibration

We confirm that in the context of ACTRIS Transnational Access the following inter-comparison was performed at the **European Center for Aerosol Calibration (ECAC)**:

Project number: **OCEC-2021-1**

Inter-comparison/Calibration date: **March 2021**

- ☒ Instrument: **Sunset Lab OCEC analyser** (serial number 461-230)
☐ Laboratory: *(institute)*

The Inter-comparison/Calibration was:

- ☒ **Successful**
☐ Unsuccessful
☐ Another calibration is recommended (see calibration report).
The PI has to submit a new TNA application.

Statistics for data submission to EBAS

Statistics	QA_bias %	QA_variability %	Systematic
TC	-3.3%	6.5%	low
OC	-3.1%	7.4%	low
EC	-4.7%	1.7%	low

Number of RWD (Researcher Working Days) : 3

Ispira, 09.06.2021

p.p. J.P. Putaud

Signature

Head of ERLAP, or ACMCC