



European Centre for Aerosol Calibration
Administrating Office
Leibniz Institute for Tropospheric Research
Permoserstr. 15
04318 Leipzig

PONT, Veronique/DIAS-ALVES , Maria
LAERO
14 avenue Edouard Belin, 31400 Toulouse, France

TNA – ECAC Confirmation of Intercomparison/Calibration

We confirm that in the context of ACTRIS Transnational Access the following inter-comparison was performed at the **European Center for Aerosol Calibration (ECAC)**:

Project number: **OCEC-2022-1**

Inter-comparison/Calibration date: **March 2022**

- Instrument: **Sunset Lab OCEC analyser** (serial number **461-230**)
 Laboratory: *(institute)*

The Inter-comparison/Calibration was:

- Successful**
 Unsuccessful
 Another calibration is recommended (see calibration report).
The PI has to submit a new TNA application.

Statistics for data submission to EBAS

Statistics	QA_bias %	QA_variability %	Systematic
TC	-8.3%	5.8%	low
OC	-10.8%	6.4%	low
EC	17.2%	16.4%	no

Number of RWD (Researcher Working Days) : 3

Ispra, 03.06.2022
p.p. J.P. Putaud
Signature
Head of ERLAP, or ACMCC