



European Centre for Aerosol Calibration  
Administrating Office  
Leibniz Institute for Tropospheric Research  
Permoserstr. 15  
04318 Leipzig

CORCUERA, Beatriz Nuñez/GARCIA, Saul  
Instituto de Salud Carlos III  
Crt. Majadahonda-Pozuelo km 2.2, 28222 MADRID, Spain

### TNA – ECAC Confirmation of Intercomparison/Calibration

We confirm that in the context of ACTRIS Transnational Access the following inter-comparison was performed at the **European Center for Aerosol Calibration (ECAC)**:

Project number: **OCEC-2022-1**

Inter-comparison/Calibration date: **March 2022**

- Instrument: **Sunset Lab OCEC analyser** (serial number 5L, 1971654834)  
 Laboratory: (*institute*)

The Inter-comparison/Calibration was:

- Successful**  
 Unsuccessful  
 Another calibration is recommended (see calibration report).  
The PI has to submit a new TNA application.

Statistics for data submission to EBAS

Statistics	QA_bias %	QA_variability %	Systematic
TC	2.7%	7.2%	high
OC	5.9%	7.0%	high
EC	-20.2%	15.2%	no

Number of RWD (Researcher Working Days) : 3

Ispra, 03.06.2022  
p.p. J.P. Putaud  
Signature  
Head of ERLAP, or ACMCC