



CPC Model: TSI CPC 3010

CPC Serial Number SN2426

Customer: TROPOS internal

Description: Calibration of a Condensation Particle Counter

Date of Calibration: April 8, 2022

Summary of Intercomparison:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 95% efficiency at 40 nm. The Dp50 is at 9.6 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Certificate / Reference: WCCAP


Date of issue: April 8, 2022 Signature

Reviewed by: TROPOS Name: Dr. Thomas Tuch

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de










	Unit	Status	
Participant		RadonTSMPS	
CPC Model		TSI 3010 SN 2426	
Firmware			
Manufacture date			
Last service date			
Arrival date			
Software Version		Labcount 2021	
Saturator Temperature	°C		
Condenser Temperature	°C		
Optics Temperature	°C		
Cabinet Temperature	°C		
Ambient Pressure	mbar	998	
Vaccuum Pressure	kPa		
Inlet Pressure	kPa		
Critical Orifice Pressure	kPa	Needle valve	
Aerosol Nozzle Pressure	kPa		
Laser Current	mA		
Liquid Level		full	
Aerosol Flow	L/min	0.988	
Zero	avg 10 min	0	
Physical inspection		ok	
Functional test		ok	

Table 1. Diagnostic information of candidate









Aerosol electrometer	TSI Electrometer Model 3068, SN 70838596	
Particles and gases used for calibration	silver particles and nitrogen	
Method of particle generation	tube furnace generator	
Electrometer calibration certificate	September, 2021, calibrated at PTB	
	Braunschweig	
Corrections of electrometer (i.e.	Within tolerance range (+/-2%);	
differing flow rate)	reference: 4.0 l/min, measured: 4.0 l/min	
Logging software	LabView 2021; National Instruments;	
	Program "LabCount.vi"	
Uncertainty in measured flow rate	3%	
Flowmeter used	Gilian Gilibrator 3; SN 21181001005, 2021	
Lab Tempreature and Pressure	23.0°C, 1013 mbar	

Table 2. Calibration and laboratory conditions


		Pulse Output	
Diameter	EL 3068B	Concentration	Efficiency
	(#/cm3)	(#/cm3)	(μ)
5	1366.33	0.17	0.00
6	1364.67	29.78	0.02
7	1435.94	178.72	0.12
8	1655.43	443.33	0.27
9	1807.45	745.76	0.41
10	1517.04	801.73	0.53
11	1272.63	801.95	0.63
14	1931.86	1670.29	0.86
20	1734.68	1646.04	0.95
30	1791.19	1706.47	0.95
40	1639.75	1568.70	0.96

Table 3. Efficiency of candidate CPC per diameter against the Electrometer

Leibniz-Institut für Troposphärenforschung e.V. Telefon: +49 341 2717-7060 Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de Commerzbank Leipzig KTO 102 14 50 BLZ 860 400 00 IBAN: DE77 8604 0000 0102 1450 00 SWIFT CODE: COBADEFF 860












Fig. 1. Counting efficiency of candidate CPC against aerosol electrometer 3068 SN 70838596; silver particles between 5 nm and 40 nm were used for calibration. Dp50 = 9.6 nm

Date of issue: April 8, 2022

Reference: TSI electrometer, model 3068, SN 70838596

Reviewed: TROPOS/Dr. Thomas Tuch

