

# Instrument Inter-Comparison Report

---

Instrument	
Type	MAAP model 5012
Serial Number	118
Institution	JRC Atmosphere, Ispra
Contact	Sebastiao Martins dos Santos

Instrument inter-comparison	
Organization	Leibniz Institute for Tropospheric Research (TROPOS) World Calibration Centre for Aerosol Physics (WCCAP)
Contact	Thomas Müller, <a href="mailto:thomas.mueller@tropos.de">thomas.mueller@tropos.de</a> Alfred Wiedensohler, <a href="mailto:ali@tropos.de">ali@tropos.de</a>
Workshop, etc.	WCCAP-2015-1, Absorption Photometer Workshop, 21-25 Sep. 2015

Report	
Status	final
Date	2015-10-20

## 1. Instrument inter-comparison summary

**Flow calibration:** The flow of the instrument agreed to the flow measured with a reference flow meter Gilibrator). The instrument flow was 0.25 % too low resulting in lower eBC concentrations. Correction of the flow error was included in the data evaluation.

**Noise.** The noise level of the instrument was little higher than expected from the MAAP specification sheet. The average noise ( $1\sigma$ ) was 80.7 ng/m<sup>3</sup> for 1 minute averaging time. For unknown reasons the noise periodically increases by a factor of about 2. The noise does not influence long term averages.

**Comparison to reference MAAP:** BC concentrations are about 2.3% higher than BC concentrations from the 'reference' MAAP.

**Comparison to reference absorption:** Absorption coefficients derived from MAAP are about 6 % higher than absorption coefficients from the Multiwavelength Absorption Reference setup.

**Flow regulation:** The flow regulation of the instrument is unstable for flows below instrument shows problems with flows lower than 700 l/hour. The reason is unclear.

**Recommendations:** None

**Overall assessment:** The instrument meets the requirements.

## 2. Instrument configuration

Configuration parameters (Print format 8)			
THERMO SCIENTIFIC	MAAP v1.32	SERIAL NUMBER 118	15-09-21
-----			
SIGMA BC:	6.6 m2/g		
AIR FLOW:	800		
STORE AVERAGES:	10 min		
VOLUME REFERENCE	STANDARD TEMPERATURE		
STANDARD TEMPERATURE	25 _C		
PRINTFORMAT:	COM2 8		
PRINTCYCLE:	1 min		
BAUDRATE:	Bd COM1 9600		
BAUDRATE:	Bd COM2 9600		
DEVICE-ADDRESS:	0		
FILTER CHANGE			
TRANSM. <	% 50		
CYCLE	h 0		
HOURL:	0		
CALIBRATION OF SENS.			
T1 T2 T3 T4 P1 P2 P3			
-39 -24 -82 67 141 -245 -234			
AIR FLOW	100.0		
HEATER PARAMETERS			
Diff. T2-T1 nominal	0 _C		
Max. Heating Temp.	45 _C		
Min. Heating Power	10 %		
ANALOG OUTPUTS			
OUTPUT ZERO:	0mA		
CBC	0 10		
MBC	0 2400		
GESYTEC-PROTOKOL			
STATUS VERSION	STANDARD		
NUMBER OF VARIABLES	7		
CBC REFLMBC Q-OPQ-N T1 P1			
END			

## 3. Data Processing

Equivalent black carbon concentrations reported by instruments were corrected for flow deviations and adjusted to standard temperature and pressure conditions (T=0°C, P=1013.25 hPa) by

$$[BC] = [BC_{instr}] \times F_{flow} \times F_{STP}$$

For details read Appendix A.

Conversion between the eBC concentrations and the absorption coefficient is done by

$$b_{abs}[1/Mm] = eBc[\mu g/cm] \times Sigma \times 1.05 ,$$

with the *mass absorption cross section*  $MAC=6.6 \text{ m}^2/\text{g}$ . During the RAOS (Sheridan et al. 2005) experiment the MAAP was compared to a reference absorption at the wavelength 670 nm, but the true wavelength of MAAP is 637 nm. The factor compensates the resulting error in the absorption (Mueller et al. 2010).

#### 4. Technical checks

**Table: Flow check**

Correction factors  $F_{flow}$  and  $F_{STP}$  for correcting eBC concentrations.  $F_{flow}$  corrects inlet flow errors.  $F_{STP}$  adjusting concentrations to STP conditions (0°C, 1013.25 hPa).

Date	System Flow			Reference flow			Flow correction factor <sup>1</sup>	STP correction factor <sup>1</sup>
				Reference flow meter: Gilibrator ‘TROPOS-T’				
	Mass flow	Volume reference		Volume flow	Ambient <i>T</i> and <i>P</i>			
<i>Q</i> <sub>MAAP</sub> [slpm]	<i>T</i> <sub>0,MAAP</sub> [°C]	<i>P</i> <sub>0,MAAP</sub> [hPa]	<i>Q</i> [lpm]	<i>T</i> [°C]	<i>P</i> [hPa]	<i>F</i> <sub>flow</sub>	<i>F</i> <sub>STP</sub>	
21. Sep	13.33	25.0	1013	13.41	24	2001	1.0025	1.092

**Table: Sample spot**

Date	Spot appearance	Spot size correction factor
21. Sep	Well defined spot, spot size not measured	1.0 <sup>1</sup>

**Table: Instrumental Noise**


Noise in units of eBC concentration measured with filtered air.

Date	Avg. time	Wave-length [nm]	Num data points	Median [ng]	10 <sup>th</sup> percentile [ng]	90 <sup>th</sup> percentile [ng]	Mean [ng]	Standard deviation [ng]	Error of the mean [ng]
Sep. 22	1 min	637	102	0.0	-58.0	72.0	2.7	80.7	8.0

#### 5. Comparison to reference instruments

The reference MAAP (SN504) was not available due to an instrumental error and was replaced by another MAAP (SN32). MAAP-SN32 was inter-compared before the workshop to two other MAAPs. The three instruments agreed within 5% and the noise level of MAAP-SN32 was in agreement with the instrumental specifications.

<sup>1</sup> See appendix A for calculations of flow, STP and spot size corrections factors.




**Table: Comparison of MAAPs**

Correlation of eBC from MAAP (SN118) and MAAP (SN032)

Slope	1.023 ± 0.011
R <sup>2</sup>	0.47

## Comparison to the Multi-Wavelength absorption reference



**Table: Correlation of absorption coefficients from MAAP (SN118) and the Multi-wavelength absorption reference (interpolated to 637 nm).**

Slope	1.061 $\pm 0.018$
R <sup>2</sup>	0.27

## Appendix A: Instrument corrections

Necessary corrections to all instruments are flow and spot size correction and conversion of concentrations and absorption coefficients to STP conditions. BC concentrations from individual instruments  $[BC_{instr}]$  were corrected by:

$$[BC] = [BC_{instr}] \times F_{flow} \times F_{spot} \times F_{STP}$$

- a) The Flow correction factor for compensating calibration errors of the instrument flow meter and is defined by:

$$F_{flow} = \frac{Q_{instr} [slpm]}{Q_{ref} [lpm]} \times \frac{T_{ref} [K]}{T_{0,instr} [K]} \times \frac{P_{0,instr} [hPa]}{P_{ref} [hPa]}$$

where  $Q_{instr.}$  and  $Q_{ref}$  are the flows measured with the instrument and determined with a reference volume flow meter, respectively. The flow of the volume flow meter is converted using the temperature  $T_{ref}$  and pressure  $P_{ref}$ , which are typically the ambient or room temperature or pressure near the reference flow meter. Also the standard temperature  $T_{0,instr}$  and standard pressure  $P_{0,instr}$  of the instrument have to be considered.

- b) The adjustment of instrument flow to standard temperature and pressure (STP) is done by

$$F_{STP} = \frac{T_{0,instr.} + 273}{T_0 + 273} \times \frac{P_0}{P_{0,instr.}}$$

- c) whereas  $T_{0,instr}$  and  $P_{0,instr.}$  are the standard temperature and pressure of individual instrument. For ACTRIS workshops STP is defined to be  $T_0=0^\circ\text{C}$  and  $P_0=1013.25$  hPa.
- d) The spot size correction factor  $F_{spot}$  compensates for systematic deviations of sample spot sizes and is defined by

$$F_{spot} = \frac{A_{meas}}{A_{instr}}$$

where  $A_{instr.}$  and  $A_{meas}$  are the instrument nominal and the measured spot area, respectively.

## References

Sheridan, P. J., et al. (2005). "The Reno Aerosol Optics Study: An evaluation of aerosol absorption measurement methods." Aerosol Science and Technology **39**(1): 1-16.

Müller, T., et al. (2011). "Characterization and intercomparison of aerosol absorption photometers: result of two intercomparison workshops." Atmospheric Measurement Techniques **4**(2): 245-268.