




Leibniz Institute for
Tropospheric Research



Intercomparison of Absorption Photometers

Project No.: AP-2017-2-4

Location of the quality assurance: TROPOS, lab 121

Date: 18 September, 2017

| Principal Investigator | Home Institution | Participant | Instrument |
|------------------------|--------------------------------------|-------------|-------------------|
| A. Eija | Finish Meteorological Institue (FMI) | J. Backmann | MAAP, SN 42545/15 |

1. Intercomparison summary

Flow calibration: The flow meter of the instrument is set to report flow for conditions of 21.11°C and 1013.25 hPa. The flow was 9.3% too low compared to reference flow meter (Gilibrator). Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

Noise: The noise level of the instrument was little higher than expected from the MAAP specification sheet. The average noise (1σ) was $30.8 \text{ ng}\cdot\text{m}^{-3}$ for 1 min averaging time.

Inspection: Measurement cell was dirty and had to be cleaned. Seal ring at the inlet had to be replaced. The sample spots showed well defined, sharp edges.

Comparison to a reference MAAP: BC concentrations are about 17% higher than BC concentrations from reference MAAP.

Comparison to reference absorption: The absorption coefficients derived from MAAP are 31% higher than absorption coefficients from the multi-wavelength absorption reference setup. The uncertainty of the reference absorption for the present concentrations is about 10% to 15%.

Recommendations: None.

Overall assessment: The instrument meets the requirements.

2. Details

Configuration parameters

SIGMA BC: 6.6 m2/g

AIR FLOW: 600

STORE AVERAGES: 1 min

VOLUME REFERENCE STANDARD TEMPERATURE

STANDARD TEMPERATURE 0_C

```
PRINTFORMAT:  COM1  5
```

PRINTCYCLE: 1 min

BAUDRATE: Bd COM1 9600

BAUDRATE: Bd COM2 9600

DEVICE-ADDRESS: 1

FILTER CHANGE

TRANSM. < % 50

CYCLE h 0

HOUR: 0

CALIBRATION OF SENS.

P1,SP P1,Z P2,SP P2,Z P3,Z T1,Z T2,Z T3,Z T4,Z

-22 -4 -36 65 -334 -200 299

| | |
|----------|------|
| AIR FLOW | 92.6 |
|----------|------|

ANALOG OUTPUTS

OUTPUT ZERO: 4mA

| | | |
|-----|---|----|
| CBC | 0 | 10 |
|-----|---|----|

MBC 0 2400

Q-OP 0 1000

T1 -20 40

T2 -20 40

P3 900 1100


GESYTEC-PROTOKOL

| STATUS | VERSION | STANDARD |
|--------|---------|----------|
|--------|---------|----------|

NUMBER OF VARIABLES 1

CBC

| Flow check | | | | | | | | |
|---|----------------------|----------------------|-----------------------|---|---------------------|--------------|--|---|
| ¹ Correction factors F_{flow} and F_{STP} for correcting eBC concentrations. F_{flow} corrects for inlet flow errors considering leakage. F_{STP} is used to adjust concentrations to STP conditions (0°C, 1013.25 hPa). | | | | | | | | |
| Date | System Flow | | | Reference flow | | | Flow correction factor ¹ Fehler! Textmarke nicht definiert. | STP correction factor ¹ Fehler! Textmarke nicht definiert. |
| | | | | Reference flow meter: Gilibrator ‘TROPOS-T’ | | | | |
| | Mass flow | Volume reference | | Volume flow | Ambient T and P | | | |
| | Q_{MAAP} [slpm] | $T_{0,MAAP}$ [°C] | $P_{0,MAAP}$ [hPa] | Q [lpm] | T [°C] | P [hPa] | F_{flow} | F_{STP} |
| 2017-09-06 | 10 | 21.11 | 1013.25 | 9.283 | 20 | 995 | 1.093 | 1.077 |

| Spot size check | | | |
|--|--------------------------------------|---|-------------------|
| Correction factor for spot sizes F_{spot} . | | | |
| Date | Nominal spot size [cm ²] | Measured spot size [mm ²] | F _{spot} |
| 2017-09-06 | 0.785 | Well defined spot, spot size not measured | 1.0 |
|  | | | |

Cell inspections

Cell was dirty, coated with a significant dust layer.

**Instrumental Noise**

Noise in units of eBC concentration measured with filtered air.

| Date | Avg. time | Wave-length [nm] | Num data points | Median [ng] | 10 th percentile [ng/m ³] | 90 th percentile [ng/m ³] | Mean [ng/m ³] | Standard deviation [ng/m ³] | Error of the mean [ng/m ³] |
|------------|-----------|------------------|-----------------|-------------|--|--|---------------------------|---|--|
| 2016-09-30 | 1 min | 637 | 466 | 0.0 | -37.5 | 26.5 | -3.56 | 30.81 | 1.43 |

| Comparison to reference MAAP Correlation of eBC from MAAP (SN 42545/15) and the reference MAAP (SN 504). | |
|--|-------------|
| Slope | 1.170±0.006 |
| R ² | 0.978 |




Figure: Comparison of eBC concentrations from MAAP SN 42545/15 and MAAP SN-504.

Comparison to multi-wavelength absorption reference
 Correlation of absorption coefficients from MAAP (SN 42545/15) and the multi-wavelength absorption reference

| | |
|----------------|-------------|
| Slope | 1.316±0.012 |
| R ² | 0.943 |




Figure: Comparison of absorption coefficients from MAAP SN 42545/15 and absorption reference.