

## Intercomparison of Absorption photometer

Project No.: AP-2017-4-3

### Basic informations:

**Location of the quality assurance:** TROPOS, lab 121

**Date:** 10 November, 2017

Principal Investigator	Home Institution	Participant	Instrument
G. Mocnik	Aerosol d.o.o.	A. Gregoric	AE33, SN #S05-00511

### 1 Intercomparison summary

**Status on arrival:** No issues due to transportation or other damages.

**Flow calibration:** The flow meter of the instrument is set to report flow for conditions of 21.11°C and 1013.25 hPa. The flow was 0.5% too high compared to reference flow meter (Gilibrator). Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

**Noise:** The noise level of the instrument is in the normal range. The average noise ( $1\sigma$ ) for the all wavelengths was less equal 86 ng·m<sup>-3</sup> for one minute averaging time. The background level was acceptable with deviations of less equal 19 ng·m<sup>-3</sup> for all wavelengths.

**Inspection:** The instrument was new without any contamination.

**Comparison to reference MAAP:** BC concentrations at 660 nm (BC5) of AE33 are 56.8% higher than BC concentrations from a reference MAAP.

**Comparison to reference AE33:** The deviations of BC concentrations relative to the reference AE33 are in the range of 29.4 to 45.6%.

**Comparison to reference absorption:** The absorption coefficients derived from AE33 at 660 nm (BC5) are 3.1% lower than absorption coefficients from the multi-wavelength absorption reference setup.

**Recommendations:** This was a test measurement with a new filter tape. The systematic deviations could be caused by this fact.

**Overall assessment:** In general, the instrument does not meet the requirements.

## 2 Details


Configuration parameters
<pre> &lt;data&gt;   &lt;name&gt;Aethalometer&lt;/name&gt;   &lt;manufacturer&gt;Magee Scientific&lt;/manufacturer&gt;   &lt;!-- Instrument serial number --&gt;   &lt;SerialNumber&gt;AE33-S05-00511&lt;/SerialNumber&gt;   &lt;!--Model number--&gt;   &lt;ModelNo&gt;AE33&lt;/ModelNo&gt;   &lt;!--Language used for all text in AE software!--&gt;   &lt;language&gt;EN&lt;/language&gt;   &lt;!-- Number of channels, 7 wavelenghts (from IR to UV)--&gt;   &lt;NoOfChannels&gt;7&lt;/NoOfChannels&gt;   &lt;About&gt;0&lt;/About&gt;   &lt;SetupStartTime&gt;2016/10/20 14:47:22&lt;/SetupStartTime&gt;   &lt;SetupEndTime&gt;2016/12/02 17:20:11&lt;/SetupEndTime&gt;   &lt;DateFormat&gt;1&lt;/DateFormat&gt;   &lt;MeasureTimeStamp&gt;1&lt;/MeasureTimeStamp&gt;   &lt;!-- Preset value for pump--&gt;   &lt;PumpPresetValue&gt;589&lt;/PumpPresetValue&gt;   &lt;!-- Set Flow in mlpm --&gt;   &lt;FlowSet&gt;5000&lt;/FlowSet&gt;   &lt;!-- TimeBase interval; can be 1, 60 seconds --&gt;   &lt;TimeBase&gt;60&lt;/TimeBase&gt;   &lt;!-- sigma value for channel 1--&gt;   &lt;SG1&gt;18.47&lt;/SG1&gt;   &lt;!-- sigma value for channel 2--&gt;   &lt;SG2&gt;14.54&lt;/SG2&gt;   &lt;!-- sigma value for channel 3--&gt;   &lt;SG3&gt;13.14&lt;/SG3&gt;   &lt;!-- sigma value for channel 4--&gt;   &lt;SG4&gt;11.58&lt;/SG4&gt;   &lt;!-- sigma value for channel 5--&gt;   &lt;SG5&gt;10.35&lt;/SG5&gt;   &lt;!-- sigma value for channel 6--&gt;   &lt;SG6&gt;7.77&lt;/SG6&gt;   &lt;!-- sigma value for channel 7--&gt;   &lt;SG7&gt;7.19&lt;/SG7&gt;   &lt;!-- Spot size in cm2--&gt;   &lt;Area&gt;0.785&lt;/Area&gt;   &lt;!-- Maximum attenuation before tape advance--&gt;   &lt;AtnMAX&gt;120&lt;/AtnMAX&gt;   &lt;!-- Condition when Tape Advance starts; 1 - ATNmax, 2 - time interval (every n-hours), 3 - certain time of day --&gt;   &lt;TAtype&gt;1&lt;/TAtype&gt;   &lt;!-- TapeAdvanceInterval is unit in hours between 2 tape advance --&gt;   &lt;TapeAdvanceInterval&gt;12&lt;/TapeAdvanceInterval&gt;   &lt;!-- TapeAdvanceTime is time of next tape advance occurence! --&gt;   &lt;TapeAdvanceTime&gt;2015/01/01 00:00:00&lt;/TapeAdvanceTime&gt;   &lt;!-- TapeAdvanceCount is overall number of TA counts! --&gt;   &lt;TapeAdvanceCount&gt;34&lt;/TapeAdvanceCount&gt;   &lt;!-- WarmUpInterval is time (in minutes) after TA of Clean Air flow--&gt;   &lt;WarmUpInterval&gt;3&lt;/WarmUpInterval&gt;   &lt;!-- Flow calculation parameters --&gt;   &lt;FlowFormulaA0&gt;-2204.955&lt;/FlowFormulaA0&gt;   &lt;FlowFormulaA1&gt;-2571.604&lt;/FlowFormulaA1&gt;   &lt;FlowFormulaA2&gt;-3000&lt;/FlowFormulaA2&gt;   &lt;FlowFormulaB0&gt;12.36972&lt;/FlowFormulaB0&gt;   &lt;FlowFormulaB1&gt;12.76701&lt;/FlowFormulaB1&gt;   &lt;FlowFormulaB2&gt;16&lt;/FlowFormulaB2&gt;   &lt;FlowFormulaC0&gt;-0.0001852726&lt;/FlowFormulaC0&gt;   &lt;FlowFormulaC1&gt;-0.0001894063&lt;/FlowFormulaC1&gt;   &lt;FlowFormulaC2&gt;-0.003&lt;/FlowFormulaC2&gt;   &lt;FlowFormulaD&gt;179.7361&lt;/FlowFormulaD&gt;   &lt;FlowFormulaE&gt;0.08127239&lt;/FlowFormulaE&gt;   &lt;FlowFormulaF&gt;1.01648E-07&lt;/FlowFormulaF&gt;   &lt;!-- Tape offset--&gt;   &lt;!-- TapeOffset 0-not set yet! 1-set tapeleft and right offset are valid --&gt;   &lt;TapeOffsetValid&gt;1&lt;/TapeOffsetValid&gt;   &lt;TapeRightFormulaK&gt;1.133987&lt;/TapeRightFormulaK&gt;   &lt;TapeRightFormulaN&gt;-28.72876&lt;/TapeRightFormulaN&gt; </pre>

```
<TapeLeftFormulaK>1.159609</TapeLeftFormulaK>
<TapeLeftFormulaN>-41.00325</TapeLeftFormulaN>
<!-- TapeAdvanceLength adjustment-->
<TapeAdvanceAdjust>15</TapeAdvanceAdjust>
<!-- Compensation algorithm-->
<Zeta>0.025</Zeta>
<C>1.57</C>
<ATNf1>10</ATNf1>
<ATNf2>30</ATNf2>
<Kmax>0.015</Kmax>
<Kmin>-0.005</Kmin>
<Aff>1</Aff>
<Abb>2</Abb>
<k0>0.006971313</k0>
<k1>0.007603685</k1>
<k2>0.008135425</k2>
<k3>0.0086183</k3>
<k4>0.009270029</k4>
<k5>0.01049349</k5>
<k6>0.01071778</k6>
<!-- Flow reporting standard-->
<FlowRepStd>0</FlowRepStd>
<!-- External Pressure -->
<P>101325</P>
<!-- External Temperature -->
<T>21.11</T>
<!-- External device on COM1-->
<Device1>0</Device1>
<!-- External device on COM2-->
<Device2>0</Device2>
<!-- External device on COM3-->
<Device3>0</Device3>
<!-- Network connection-->
<IPaddress>
</IPaddress>
<AutoConnect>0</AutoConnect>
<!-- Auto test enabled 0-NO, 1-YES -->
<AutoTestEnabled>1</AutoTestEnabled>
<!-- Auto test type - weekly = 0, monthly = 1-->
<AutoTestType>0</AutoTestType>
<!-- Auto test day 1-SUN, 2-MON, 3-TUE, 4-WED, 5-THU, 6-FRI, 7-SAT-->
<AutoTestDay>5</AutoTestDay>
<!-- Auto test Time-->
<AutoTestTime>2015/01/01 00:00:00</AutoTestTime>
<!-- Dispaly settings - 0 - ON, 1 - Screen Saver, 2 - Auto OFF-->
<Display>1</Display>
<!-- Timezone settings, DST -->
<TimeZone>Coordinated Universal Time</TimeZone>
<DaylightSavingTime>0</DaylightSavingTime>
<HomeInfo>0</HomeInfo>
<ExternalID>1</ExternalID>
<BHparamID>1</BHparamID>
</data>
```

<b>Flow check</b>								
<sup>1</sup> Correction factors Fflow and FSTP for correcting eBC concentrations. Fflow corrects for inlet flow errors considering leakage. FSTP is used to adjust concentrations to STP conditions (0°C, 1013.25 hPa).								
Date	System Flow			Reference Flow			Flow correction factor <sup>1</sup>	STP correction factor <sup>1</sup>
	Mass flow	Volume reference		Volume flow	Ambient T and p			
	$Q_{MAAP}$ [slpm]	$T_{0,MAAP}$ [°C]	$p_{0,MAAP}$ [hPa]	Q [lpm]	T [°C]	P [hPa]	$F_{flow}$	$F_{STP}$
2017-11-04	5.0	21.11	1013.25	5.091	25	1013.25	0.995	1.077

<b>Spot size check</b>			
Correction factor for spot sizes $F_{spot}$ .			
Date	Nominal spot size [cm <sup>2</sup> ]	Measured spot size [cm <sup>2</sup> ]	$F_{spot}$
2017-11-04	0.785	Well defined spot, spot size not measured	1.0



<b>Instrumental Noise</b>									
Noise in units of eBC concentration measured with filtered air.									
Date	Avg. time	Wave-length [nm]	Num data points	Median [ng]	10 <sup>th</sup> percentil e [ng/m <sup>3</sup> ]	90 <sup>th</sup> percentil e [ng/m <sup>3</sup> ]	Mean [ng/m <sup>3</sup> ]	Standard deviation [ng/m <sup>3</sup> ]	Error of the mean [ng/m <sup>3</sup> ]
2017-11-06	1 min	370	121	19	-16	49	23	65	6
		470	121	-5	-27	18	3	78	7
		520	121	-2	-32	26	5	81	7
		590	121	-8	-41	33	2	86	8
		660	121	-4	-34	28	5	77	7
		880	121	-13	-42	20	-6	61	6
		950	121	-13	-43	23	-4	61	6



**Comparison to reference AE33**

Correlation of eBC coefficients from AE33 (S05-00511) and reference AE33.

wavelength	Slope	R <sup>2</sup>
370	1.294±0.006	0.986
470	1.424±0.006	0.988
520	1.425±0.006	0.988
590	1.414±0.006	0.988
660	1.375±0.006	0.988
880	1.431±0.007	0.985
950	1.456±0.007	0.985



### Comparison to multi-wavelength absorption

Correlation of absorption from AE33 (S05-00511) and the multi-wavelength absorption reference at 660 nm.

Slope	$1.031 \pm 0.009$
$R^2$	0.955

