



Leibniz Institute for
Tropospheric Research



Intercomparison of Absorption photometer

Project No.: AP-2018-1-1

Basic informations:

Location of the quality assurance: TROPOS, lab 121

Date: 19 March 2018

Principal Investigator	Home Institution	Participant	Instrument
B. Bergmans	ISSeP	B. Bergmans	AE22 (Fehler! Verweisquelle konnte nicht gefunden werden.)

1 Intercomparison summary

Status on arrival: Instrument was not wrapped tight enough. An broken inlet connector was plugged in during transport.

Flow calibration: The flow meter of the instrument is set to report flow for conditions of 20°C and 1013 hPa. The flow was 1.2% too low compared to reference flow meter (Gilibrator). Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

Noise: The noise level of the instrument is in the normal range. The average noise (1σ) for the all wavelengths was less equal 80 $\text{ng}\cdot\text{m}^{-3}$ for one minute averaging time. The background level was acceptable with deviations of less equal 49 $\text{ng}\cdot\text{m}^{-3}$ for all wavelengths.

Inspection: The instrument was almost clean. The measuring head was not removable properly, therefor not checked.

Comparison to reference MAAP: BC concentrations at 880 nm of AE22 are 33.1% lower than BC concentrations from a reference MAAP.

Comparison to reference AE33: BC concentrations at 880 nm of AE22 are 29.9% lower than BC concentrations from a reference AE33.

Recommendations: The deviations appear to be systematic. The instrument requires a thorough inspection.

Overall assessment: In general, the instrument does not meet the requirements.

2 Details

Configuration parameters
<pre>--- AE-SETUP.TXT --- Created : 16.03.2018 10:02:01 . Instrument serial number: 1111 Software version: AF985d7 Instrument type (0..U (1X), 1..UV+LED (2X), 2..7xLED (3X)): 1 Instrument Chassis : Stationary Smoothing factor : 0 Selected Pump Flow : 4.0 LPM Flow scale factor : 1.84 LPM/V Flow zero : .026V Date format (0=US, 1=EU): 1 Tape saver: 0 Spots per advance: 1 Filter change interval: 0 Maximum attenuation: 150 Over old data: 1 Warm up wait: 0 Spot size: Extended Range MeanRatio: .78 BC Unit (0..ng, 1..ug): 0 . Serial comm. mode (1..OFF, 2..Dataline, 3..Gesytec): 2 Serial communication parameters: Speed(bps) : 9600 Data bits : 8 Parity bits:N Stop bits : 1 . Gesytec parameters: Network Scale Factor: 10 Instrument ID for Gesytec:333 . Dataline parameters: Alarm mode (0..Analog out, 1..Alarm): 0 Alarm ON/OFF : 1 Alarm value limit: 10 Alarm channel selection (channel number): 1 . Data format (0..Extended, 1..Compressed): 0 Prepend SerNumber to dataline (0..No, 1..Yes): 0 . UV channel OFF (0..UV ch. ON, 1..UV ch. OFF): 0 . Sigma values: Sigma 1 : 16.6 Sigma 2 : 39.5 Sigma 3 : 0 Sigma 4 : 0 Sigma 5 : 0 Sigma 6 : 0 Sigma 7 : 0 Volumetric unit settings: Volumetric units (0..Standard, 1..Volumetric): 0 Air Pressure(mbars): 1013 Temperature(C): 20</pre>

Flow check


¹Correction factors F_{flow} and F_{STP} for correcting eBC concentrations. F_{flow} corrects for inlet flow errors considering leakage. F_{STP} is used to adjust concentrations to STP conditions (0°C, 1013.25 hPa).

Date	System Flow			Reference Flow			Flow correction factor ¹	STP correction factor ¹		
	Mass flow	Volume reference		Volume flow	Ambient T and p					
	Q_{AE22} [slpm]	$T_{0,\text{AE22}}$ [°C]	$p_{0,\text{AE22}}$ [hPa]	Q [lpm]	T [°C]	P [hPa]				
2018-03-20	4.0	20.0	1013	4.023	20	995	1.012	1.072		

Spot size check

Correction factor for spot sizes F_{spot} .

Date	Nominal spot size [cm ²]	Measured spot size [cm ²]	F_{spot}
2018-03-20	NA	Well defined spot, spot size not measured	1.0




Instrumental Noise Noise in units of eBC concentration measured with filtered air.									
Date	Avg. time	Wave-length [nm]	Num data points	Median [ng]	10 th percentil e [ng/m ³]	90 th percentil e [ng/m ³]	Mean [ng/m ³]	Standard deviation [ng/m ³]	Error of the mean [ng/m ³]
2018-03-20	5 min	370	19	-44	-90	-35	-53	31	7
		880	19	-49	-216	-27	-83	80	18

Comparison to reference MAAP

Correlation of eBC from AE22 (Fehler! Verweisquelle konnte nicht gefunden werden.) and the reference MAAP (SN 504).

Slope	0.669 ± 0.007
R^2	0.991



Comparison to reference AE33

Correlation of eBC coefficients from AE22 (Fehler! Verweisquelle konnte nicht gefunden werden.) and reference AE33.

wavelength	Slope	R^2
370	0.618 ± 0.009	0.985
880	0.701 ± 0.007	0.992

