



## Intercomparison of absorption photometer Project No.: AP-2019-2-4

### Basic informations:

Location of the quality assurance: TROPOS, Lab 121  
Date: 3 June - 7 June 2019

Principal Investigator	Home Institution	Participant	Instrument
H. Servomaa	FMI	H. Servomaa	483:0403

---

## 1 Intercomparison summary

### Status on arrival

No issues due to transportation or other damages.

### Flow calibration

The flow meter of the instrument is set to report flow for conditions of 20 °C and 1013.25 hPa. The flow was 4.4% too low compared to reference flow meter (TSI 4100). Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

### Noise

The noise level of the instrument is in the normal range. The average noise ( $1\sigma$ ) for the all wavelengths was less equal  $11 \text{ ng m}^{-3}$  for two minute averaging time. The background level was acceptable with deviations of less equal  $4 \text{ ng m}^{-3}$  for all wavelengths.

### **Inspection**

The measuring cell was dirty with a larger fragment that led to shaded areas at the spot. The measuring cell was cleaned.

### **Comparison to reference MAAP**

BC concentrations at 880 nm (BC6) of AE31 are 19.7 % higher than BC concentrations from a reference MAAP.

### **Comparison to reference AE33**

The deviations of BC concentrations relative to the reference AE33 are in the range of  $-4.8$  to  $0.7$  %.

### **Comparison to reference absorption**

The deviations of the absorption coefficients derived from AE31 relative to the absorption coefficients from the multi-wavelength absorption reference setup are in the range of  $-13.9$  to  $-11.8$  %.

### **Recommendations**

No recommendations.

### **Overall assessment**

The instrument meets the requirements.

## 2 Details

### Configuration parameters

```

---- AE-SETUP.TXT ----
Created : 03-jun-19 10:47:49
.
Instrument serial number: 483
Software version: 984zz
Instrument type (0..U (1X), 1..UV+LED (2X), 2..7xLED (3X)): 2
Instrument Chassis : Stationary
Smoothing factor : 0
Selected Pump Flow : 4.0 LPM
Flow scale factor : 1.78 LPM/V
Flow zero : .024V
Date format (0=US, 1=EU): 0
Tape saver: 0
Spots per advance: 2
Filter change interval: 0
Maximum attenuation: 75
Over old data: 1
Warm up wait: 0
Spot size: Standard Range
MeanRatio: 1.00
BC Unit (0..ng, 1..ug): 0
.
Serial comm. mode (1..OFF, 2..Dataline, 3..Gesyttec): 2
Serial communication parameters:
  Speed(bps) : 9600
  Data bits : 8
  Parity bits:N
  Stop bits : 1
.
Gesyttec parameters:
  Network Scale Factor: 10
  Instrument ID for Gesyttec:333
.
Dataline parameters:
Alarm mode (0..Analog out, 1..Alarm): 0
Alarm ON/OFF : 1
Alarm value limit: 10
Alarm channel selection (channel number): 1
.
Data format (0..Extended, 1..Compressed): 0
.
UV channel OFF (0..UV ch. ON, 1..UV ch. OFF): 0
.
Sigma values:
  Sigma 1 : 39.5
  Sigma 2 : 31.1
  Sigma 3 : 28.1
  Sigma 4 : 24.8
  Sigma 5 : 22.2
  Sigma 6 : 16.6
  Sigma 7 : 15.4
Volumetric unit settings:
  Volumetric units (0..Standard, 1..Volumetric): 1
  Air Pressure(mbars): 1013
  Temperature(C): 20

```

## Flow check

Table 1: Correction factors  $F_{flow}$  and  $F_{STP}$  for correcting eBC concentrations.  $F_{flow}$  corrects for inlet flow errors considering leakage.  $F_{STP}$  is used to adjust concentrations to STP conditions (0 °C, 1013.25 hPa).

System flow and reference			Measured	$F_{flow}$	$F_{STP}$
$Q_{AE31}$	$T_{0,AE31}$	$p_{0,AE31}$	flow $Q$		
[slpm]	[°C]	[hPa]	[slpm]		
3.9	20	1013.25	3.75	1.044	1.073

## Spot size check

Table 2: Correction factor for spot sizes  $F_{spot}$ .

Nominal spot size	Measured spot size	$F_{spot}$
[cm <sup>2</sup> ]	[cm <sup>2</sup> ]	
-	Unsharp spot with shaded area, spot size not measured	1.0

## Instrumental Noise

Table 3: Noise parameters of AE31 (483:0403) measured with filtered air.

Wavelength	Number	Median	10th	90th	Mean	Std.	Error
[nm]	of data	[ng m <sup>-3</sup> ]	percentile	percentile	[ng m <sup>-3</sup> ]	dev.	of mean
	points		[ng m <sup>-3</sup> ]	[ng m <sup>-3</sup> ]		[ng m <sup>-3</sup> ]	[ng m <sup>-3</sup> ]
370	86	-1	-13	9	-2	9	1
470	86	-1	-11	6	-3	7	1
520	86	-2	-16	11	-2	10	1
590	86	-2	-13	9	-2	10	1
660	86	-3	-17	10	-4	10	1
880	86	-4	-16	9	-4	9	1
950	86	-3	-20	11	-4	11	1




Figure 1: New spot from AE31 (483:0403) on filter tape.

## Comparison to reference MAAP

Table 4: Correlation parameter of eBC coefficient (BC6) from AE31 (483:0403) ( $k = 0.004$ ) and reference MAAP after inspection.

Wavelength [nm]	Slope	Error	$R^2$
880	1.197	0.008	0.999




Figure 2: Correlation of eBC coefficient (BC6) from AE31 (483:0403) and reference MAAP.

## Comparison to reference AE33

Table 5: Correlation parameter of eBC coefficients from AE31 (483:0403) ( $k = 0.004$ ) and reference AE33 after inspection.

Wavelength [nm]	Slope	Error	$R^2$
370	0.978	0.016	0.992
470	0.952	0.012	0.995
520	0.98	0.012	0.996
590	0.976	0.011	0.996
660	0.979	0.01	0.997
880	1.007	0.009	0.998
950	0.961	0.009	0.997




Figure 3: Correlation of eBC coefficient (BC6) from AE31 (483:0403) and reference AE33.

## Comparison to multi-wavelength absorption

Table 6: Correlation parameter of absorption from AE31 (483:0403) ( $k = 0.004$ ,  $C_0 = 3.5$ ) and the multi-wavelength absorption reference after inspection.

Wavelength [nm]	Slope	Error	$R^2$
470	0.865	0.002	1
520	0.882	0.002	1
660	0.861	0.003	1




Figure 4: Correlation of absorption from AE31 (483:0403) and the multi-wavelength absorption reference at 660 nm.