



## Intercomparison of absorption photometer Project No.: AP-2020-2-2

### Basic informations:

Location of the quality assurance: TROPOS, Lab 121  
Date: 29 June - 03 July 2020

Principal Investigator	Home Institution	Participant	Instrument
J. S. Henzing	TNO	M. Moerman	167

---

## 1 Intercomparison summary

### Status on arrival

The filter tape was loose in the device. Apart from that, there were no further issues due to transportation or other damages.

### Flow calibration

The flow meter of the instrument is set to report flow for conditions of 0 °C and 1013.25 hPa. The flow was 23.0 % too low compared to reference flow meter (TSI 4100). Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

### Noise

The noise level of the instrument is in the normal range. The average noise ( $1\sigma$ ) was less equal  $43 \text{ ng m}^{-3}$  for one minute averaging time. The background level was acceptable

with deviations of less equal  $0 \text{ ng m}^{-3}$ .

### **Inspection**

The measuring cell was heavily contaminated with a dark, sticky substance. The cell was cleaned. The flow was recalibrated.

### **Comparison to reference MAAP**

BC concentrations of MAAP are 9.3 % higher than BC concentrations from a reference MAAP.

### **Comparison to reference absorption**

The deviations of the absorption coefficients derived from MAAP relative to the absorption coefficients from the multi-wavelength absorption reference setup is 8.0 %.

### **Recommendations**

No recommendations.

### **Overall assessment**

The instrument meets the requirements.

## 2 Details

### Configuration parameters

THERMO SCIENTIFIC		MAAP v1.32	SERIAL NUMBER 167		20-06-30	
<hr/>						
SIGMA BC:		6.6	m2/g			
AIR FLOW:		500				
STORE AVERAGES:		1	min			
VOLUME REFERENCE		OPERATING CONDITIONS				
STANDARD TEMPERATURE		0	_C			
PRINTFORMAT:		COM1	5			
PRINTCYCLE:			5 min			
BAUDRATE:		Bd	COM1	9600		
BAUDRATE:		Bd	COM2	9600		
DEVICE-ADDRESS:		0				
FILTER CHANGE						
TRANSM. <		%	30			
CYCLE		h	100			
HOUR:		24				
CALIBRATION OF SENS.						
T1	T2	T3	T4	P1	P2	P3
-15	8	-47	62	-218	-9	-174
AIR FLOW		95.1				
HEATER PARAMETERS						
Diff. T2-T1 nominal		0	_C			
Max. Heating Temp.		45	_C			
Min. Heating Power		10	%			
ANALOG OUTPUTS						
OUTPUT ZERO:		4mA				
CBC	0	10				
MBC	0	2400				
GESYTEC-PROTOKOL						
STATUS VERSION		STANDARD				
NUMBER OF VARIABLES		1				
CBC						
END						

### Flow check

Table 1: Correction factors  $F_{flow}$  and  $F_{STP}$  for correcting eBC concentrations.  $F_{flow}$  corrects for inlet flow errors considering leakage.  $F_{STP}$  is used to adjust concentrations to STP conditions (0 °C, 1013.25 hPa).

System flow and reference			Measured	$F_{flow}$	$F_{STP}$
$Q_{MAAP}$	$T_{0,MAAP}$	$p_{0,MAAP}$	flow $Q$		
[slpm]	[°C]	[hPa]	[slpm]		
7.55	0	1013.25	6.612	1.23	1

## Spot size check

Table 2: Correction factor for spot sizes  $F_{spot}$ .

Nominal spot size [cm <sup>2</sup> ]	Measured spot size [cm <sup>2</sup> ]	$F_{spot}$
2.00	Well defined spot, spot size not measured	1.0



Figure 1: New spot from MAAP (167) on filter tape.

## Instrumental Noise

Table 3: Noise parameters of MAAP (167) measured with filtered air.

Wavelength [nm]	Number of data points	Median [ng m <sup>-3</sup> ]	10th percentile [ng m <sup>-3</sup> ]	90th percentile [ng m <sup>-3</sup> ]	Mean [ng m <sup>-3</sup> ]	Std. dev. [ng m <sup>-3</sup> ]	Error of mean [ng m <sup>-3</sup> ]
660	301	0	-50	43	-2	43	3

## Comparison to reference MAAP

Table 4: Correlation parameter of eBC coefficients from MAAP (167) and reference MAAP.

Wavelength [nm]	Slope	Error	$R^2$
660	0.907	0.003	0.998




Figure 2: Correlation of eBC coefficient from MAAP (167) and reference MAAP.

## Comparison to multi-wavelength absorption

Table 5: Correlation parameter of absorption from MAAP (167) and the multi-wavelength absorption reference.

Wavelength [nm]	Slope	Error	$R^2$
637	1.08	0.004	0.999




Figure 3: Correlation of absorption from MAAP (167) and the multi-wavelength absorption reference at 660 nm.