

## Calibration workshop on Cloud Condensation Nucleus Counters

<i>Project No.:</i>	<b>CCNC-2016-1-4</b>
<i>Principal Investigator:</i>	<i>Amewu Antoinette Mensah</i>
<i>Home Institution:</i>	<i>Switzerland, ETH</i>
<i>Participant:</i>	<i>Franz Friebe, Nadine Borduas</i>
<i>Candidate:</i>	<b>CCN-100</b>
<i>Made by:</i>	DMT
<i>Counter (SN):</i>	CCN-100, SN 0507-44
<i>Software:</i>	DMT 5.0.6
<i>Location of the quality assurance:</i>	TROPOS Leipzig, lab 118
<i>Comparison period:</i>	October 24, 2016 – October 28, 2016
<i>Last Intercomparison (with Project No.):</i>	

### Summary of Intercomparison


#### *Pre-Status:*

The instrument arrived with participant. The column was wetted and a pre-status measurement was done on ambient aerosol. During the Pre-Status, the performance of the system showed relative differences of -2.0% to +7.5% compared with the TROPOS Reference Instrument SN108 for supersaturation between 0.1% and 1.0%. The system was operated with a flow of 500ml.


#### *Final Status:*

The new parameters for flow and supersaturation calibration were set. During the Final Status the performance of the system showed relative differences of -8.3% to +12.5% compared with the TROPOS Reference Instrument SN108 for supersaturation between 0.1% and 1.0%. The candidate passed the quality standards of ACTRIS and GAW.

## Laboratory setup:



**Supersaturation calibration protocol (Ammonium Sulfate Particle, size selected by TROPOS Reference MPSS No.2)**




**Figure 01:** Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on October 24<sup>rd</sup>, 2016.


**Pre-status settings:**

Date of check: 24.10.2016

<i>Calibration tab settings</i>	ETH DMT-CCN-100
<i>Sample slope</i>	64.251
<i>Sample y-intercept</i>	-146.36
<i>Sheath slope</i>	1091.6
<i>Sheath y-intercept</i>	-2486.2
<i>Temp gradiet slope</i>	17.553
<i>Temp gradient intercept</i>	1.0247

**Zero-test with filter: passed ( $< 0.02$  particles  $\text{cm}^{-3}$ )****Candidate against TROPOS CCN-100 SN 108 during the pre-status: Time Series****Figure 02:** Time series (Oct 24, 2016 06:00 pm – Oct 25, 2016 06:00 am) of the Candidate vs. TROPOS CCN-100 SN-108.

# **Candidate against TROPOS CCN-100 SN 108 during the pre-status: average over supersaturation**



**Figure 03:** Average (Oct 24, 2016 06:00 pm – Oct 25, 2016 06:00 am) of the Candidate vs. TROPOS CCN-100 SN-108.

## Flow calibration protocol (Bubble flow meter ‘Giliblator’, Gilian (Sensidyne))

	Old	New		Old	New
Sample Slope	64.251	59.639	Sheath Slope	1091.6	1121.1
Sample intercept	-146.36	-132.8	Sheath intercept	-2486.2	-2576.4

### Sample Calibration Without Sheath

Valve Set (V)	Sample Volt (V)	Sheath ' Measured Total Flow (mlpm)
2.2	2.56	20.49
2.3	2.98	44.52
2.4	3.81	93.84
2.35	3.38	70.2
2.25	2.74	29.78

### Sheath Calibration

Valve Set (V)	Sample Volt (V)	Sheath ' Measured Total Flo	Sample Calcu (mlpm)	Sheath Flow (mlpm)
2.6	2.49	2.51	249.7	15.74111
2.65	2.54	2.54	285.6	18.72306
2.75	2.63	2.6	370.4	24.09057
2.85	2.75	2.67	455	31.24725
2.95	2.86	2.75	539.4	37.80754
3.05	2.98	2.82	628.4	44.96422
3.2	3.2	2.94	778	58.0848

	input total Flow
before	503.9
after	506.4




Figure 04: Sheath flow calibration






Figure 05: Aerosol flow calibration.

## Calibration of Supersaturation- $\Delta T$ in CCN chamber

### Experiment setup:



- Solution: ammonium sulfate 0.05 mol/L.
- Particle classifier was operated in diameter-scanning mode.
- Size-resolved activation ratio of ammonium sulfate particles was measured at 6  $\Delta T$ .
- Size-resolved activation ratio curves were fitted with 2 error functions, and critical diameter was taken as the centre diameter of the second error function (Fig. 05).
- Equivalent supersaturation at each pre-selected  $\Delta T$  was derived from the fitted critical diameter based on a lookup-table according to the Standardized protocol for CCN measurements WP3-NA3 / D3.11.
- Calibration parameters was derived by a linear fit of equivalent supersaturation and  $\Delta T$  (Fig. 06).




**Figure 06:** Example for activation curve of ammonium sulfate. Size selection was done with ref2 of WCCAP. The red line gives the sum of two sigmoid fits which are fitted to the measurements data. The grey lines give the fraction of doubly charged particles and the fit corrected for the doubly charged particles. The red vertical line gives the position of the determined critical diameter.




**Table 01:** Result of the supersaturation calibration.

deltaT	SS	Dcrit	T
14.3449	0.900788698	30.48135665	29.89780867
14.3449	0.924454506	30.18477847	28.63902847
14.3449	0.921814417	30.23946728	28.64036133
10.57423	0.644749993	38.13179716	27.69563351
10.57423	0.611016347	39.45835622	27.6955341
8.06045	0.450415702	48.16204817	26.80744163
8.06045	0.443796394	48.61910194	26.81108873
5.54667	0.298278629	62.96130205	26.09650548
5.54667	0.292431067	63.7673257	26.10022458
4.28978	0.197822854	82.11867139	25.88553774
4.28978	0.198119151	82.03871692	25.88600284
4.28978	0.198880134	81.83640395	25.88633555
3.03289	0.085033256	142.4107042	25.29760202
3.03289	0.085668811	141.7202669	25.30030603




**Figure 07:** New supersaturation calibration coefficients.

## Final Status of the Candidate: time series



**Figure 08:** Time series (Oct 26, 2016 06:00 pm – Oct 27, 2016 06:00 am) of the Candidate vs. TROPOS CCN-100 SN-108.

## Final Status of the Candidate: average over supersaturation



**Figure 09:** Average of activated particles vs. supersaturation (Oct 26, 2016 06:00 pm – Oct 27, 2016 06:00 am) of the Candidate vs. TROPOS CCN-100 SN-108.