

## Intercomparison of Condensation Particle Counter

Project No.: CPC-2016-4-27

### Basic information:

<b>Principal Investigator:</b>	Alfred Wiedensohler
<b>Home Institution:</b>	Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig
<b>Participant:</b>	
<b>Instrument No.1:</b>	DE-TROPOS Wolken CPC Model 3772, SN: 3772114401
<b>Location of the quality assurance:</b>	TROPOS Leipzig, lab 130
<b>Comparison period:</b>	July 14, 2016
<b>Last Intercomparison (with Project No.):</b>	
<b>TROPOS Reference Instrument:</b>	Electrometer: TSI model 3068B #70838596, Last calibration in April 2016
<b>Additional Equipment:</b>	Bubble flow meter 'Gilibrator', Gilian (Sensidyne)

### Summary of Intercomparison:


**Final status:**

The TSI CPC Model 3772 passed the quality standards of ACTRIS and GAW.

**pre-status: 06.07.2016**

Pre CPC-Workshop: The TSI CPC Model 3772 reached 98% efficiency at 40 nm. The Dp50 is at 8.5nm. The CPC efficiency curve corresponds to the standard.

Dp in nm	counting efficiency	CPC Variable	Status
40.00	0.98	Saturator Temp	39.0
30.00	0.96	Condenser Temp	22.0
20.00	0.93	Optics Temp	40.0
15.00	0.86	Cabinet Temp	35.1
13.00	0.80	Ambient Press.	100.2
11.00	0.70	Orifice Press.	94.8
10.00	0.63	Nozzle Press.	3.6
9.00	0.57	Laser	41
8.00	0.43	Liquid level	full
7.00	0.25	Aerosol flow (l/min)	0.085
6.00	0.05		




**Figure 01:** CPC efficiency curve. Based on Electrometer TSI 3068B. Serial number: 70838596

**final-status: 14.07.2016**

1. Status of the wick: It was necessary to change the wick.
2. Status of the optics: Laser current was fine. It was necessary to clean the optics.
3. General status of the CPC: The CPC was working properly.
4. Final CPC-Workshop: The TSI CPC Model 3772 reached 100% efficiency at 40 nm. The Dp50 is at 7.8 nm. The CPC efficiency curve corresponds to the standard.

Dp in nm	counting efficiency	CPC Variable	Status
40.00	1.04	Saturator Temp	39.0
20.00	0.99	Condenser Temp	22.0
15.00	0.93	Optics Temp	40.0
11.00	0.75	Cabinet Temp	35.1
10.00	0.67	Ambient Press.	100.2
8.00	0.41	Orifice Press.	84.8
6.00	0.05	Nozzle Press.	2.6
		Laser	41
		Liquid level	full
		Aerosol flow (l/min)	1.047



**Figure 01:** CPC efficiency curve. Based on Electrometer TSI 3068B. Serial number: 70838596



**Kalibrierschein**  
Calibration Certificate

Gegenstand:  
*Object:* Aerosol Elektrometer  
ID-Nummer: TSI\_3068B\_70838596

Hersteller:  
*Manufacturer:* TSI Incorporated

Typ:  
*Type:* 3068B

Kennnummer:  
*Serial No.:* 70838596

Auftraggeber:  
*Applicant:* AG 3.23 - Herr Nowak, Tel.: 3228

Anzahl der Seiten:  
*Number of pages:* 4

Geschäftszeichen:  
*Reference No.:* PTB AG 2.11-1461323527-0327

Kalibrierzeichen:  
*Calibration mark:* 20790 PTB 16

Datum der Kalibrierung:  
*Date of calibration:* 22.04.2016

Im Auftrag:  
*On behalf of PTB:* Braunschweig, 04. Mai. 2016

Im Auftrag:  
*On behalf of PTB:*

391 008 n

Dr. B. Schumacher

Siegel  
*Seal*



C. Rohrig

Kalibrierscheine ohne Unterschrift und Siegel haben keine Gültigkeit. Dieser Kalibrierschein darf nur unverändert weiterverbreitet werden. Auszüge bedürfen der Genehmigung der Physikalisch-Technischen Bundesanstalt.

*Calibration certificates without signature and seal are not valid. This calibration certificate may not be reproduced other than in full. Extracts may be taken only with permission of the Physikalisch-Technische Bundesanstalt.*