



World Calibration Centre
for Aerosol Physics

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig



Leibniz Institute for
Tropospheric Research

CPC Model: TSI CPC 3750

CPC Serial Number: 3750184002

Customer: TSI Instruments Ltd.

Description: Calibration of a Condensation Particle Counter (CPC, Model 3750)

Date of Calibration: November 09, 2018

Certificate / Reference: WCCAP

Date of issue: November 12, 2018

Signature:

Reviewed by: **TROPOS**

Name: **Kay Weinhold**



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research

Date of arrival of instrument in calibration lab: *November 07, 2018*
Instrument: *Condensation Particle Counter*
Model and serial number of instrument: *CPC 3750 S/N 3750184002*

Result of physical inspection: *no damages*
Result of functional test: *functional test successful, no problems*

Internal parameters of instrument *nominal flow rate 1.0 l/min*

Model and identification number of aerosol electrometer: *TSI Electrometer Model 3068, S/N 70838596*

Electrometer calibration certificate: *September 05, 2018, calibrated at PTB Braunschweig*

Corrections of electrometer, for instance, differing flow rate: *Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.000 l/min*

Software for recording: *LabView 2010; National Instruments; Program „LabCount.vi“*

Date of calibration: *November 09, 2018*
Lab temperature and pressure: *23°C, 998.3 mbar*
Measured aerosol flow rate of CPC: *1.007 l/min*
Uncertainty in measured flow rate: *3%*
Flowmeter used: *Gilian Gilibrator V; S/N 1711008-S, January, 2018*

Particles and gases used for calibration: *silver particles and nitrogen*
Method of particle generation: *tube furnace generator*
Zero measurement of instrument: *0 particles/cm³ in 5 minutes*

Results (using pulse output):


Particle size (nm)	40	30	20	15	12
Number concentration (cm-3)	1130	1131	1597	1315	1360
Counting efficiency η	1.02	1.01	1.00	0.98	0.94
Particle size (nm)	10	09	08	07	06
Number concentration (cm-3)	1027	1071	1135	1294	885
Counting efficiency η	0.87	0.81	0.72	0.60	0.43
Particle size (nm)	05	40			
Number concentration (cm-3)	204	1172			
Counting efficiency η	0.17	1.01			



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research



Page 2 / 3

Fig. 1: Counting efficiency for CPC 3750 SN 3750184002 against aerosol electrometer 3068 SN 70838596; silver particles between 5 and 40 nm were used for calibration; the calculated D_{p50} is 6.41 nm.

Status information:

Status	<i>T SAT</i>	<i>T CON</i>	<i>T OPT</i>	<i>T CAB</i>	<i>P AMB</i>
from display	39.0	18.0	40.0	23.2	100.2
Status	<i>P OR</i>	<i>P NO</i>	<i>Laser</i>	<i>LV</i>	<i>flow</i>
from display	81.0	2.58	40	full	1.007

Date of issue: November 12, 2018

Reference: TSI electrometer, model 3068, SN 70838596

Reviewed: TROPOS / Kay Weinhold



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research

Page 3 / 3

Leibniz-Institut für Troposphärenforschung e.V.
Telefon: +49 341 2717-7060
Telefax: +49 341 2717-99-7060
info@tropos.de
<http://www.tropos.de>

Commerzbank Leipzig
KTO 102 14 50
BLZ 860 400 00
IBAN: DE77 8604 0000 0102 1450 00
SWIFT CODE: COBADEFF 860

Mitglied der

Leibniz-Gemeinschaft