







Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Intercomparison of Condensation Particle Counter

Project No.: CPC-2019-5-8

Principal Investigator: Christian Maier

Home Institution: ZAMG, Austria

Participant: Christian Maier and Gerhard Schauer

Candidate: CPC ENVI

Counter (SN): PALAS CPC Model ENVI #9532

Location of the quality assurance: TROPOS Leipzig, lab 130

Comparison period: October 08, 2019

Last Intercomparison (with Project No.):

TROPOS Reference Instrument: Electrometer: TSI model 3068B

#70838596, Last calibration in September 2018

Additional Equipment: Bubble flow meter 'Gilibrator', Gilian (Sensidyne)

#1711008-S, Last calibration in January 2018

Summary of Intercomparison

Status:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 100% efficiency at 40 nm. The Dp50 is at 8.03 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Page 1 / 4








for Aerosol Physics





Leibniz Institute for Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig



Leibniz-Institut für Troposphärenforschung e.V.
Telefon: +49 341 2717-7060
Telefax: +49 341 2717-99-7060

Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de Commerzbank Leipzig KTO 102 14 50 BLZ 860 400 00 IBAN: DE77 8604 0000 0102 1450 00 SWIFT CODE: COBADEFF 860 Mitglied der Leibniz-Gemeinschaft









Leibniz Institute for Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Date of arrival of instrument in calibration lab: October 08, 2019

Instrument:

Condensation Particle Counter Model and serial number of instrument: PALAS CPC ENVI S/N 9532

for Aerosol Physics

Result of physical inspection: no damages **Result of functional test:** no repair

Internal parameters of instrument nominal flow rate 0.9 l/min

Model and identification number of

aerosol electrometer: TSI Electrometer Model 3068, S/N 70838596

Electrometer calibration certificate: September 05, 2018, calibrated at PTB

Braunschweig

Corrections of electrometer, for instance,

differing flow rate:

l/min, measured: 4.00 l/min

Within tolerance range (+/-2%); reference: 4.0

Software for recording: LabView 2010; National Instruments; Program

"LabCount.vi"

Date of calibration: October 08, 2019 Lab temperature and pressure: 22.2°C, 985 mbar Measured aerosol flow rate of CPC: 0.893 l/min

Uncertainty in measured flow rate: 3%

Flowmeter used: Gilian Gilibrator V; S/N 1711008-S,

January, 2018

Particles and gases used for calibration: silver particles and nitrogen Method of particle generation: tube furnace generator **Zero measurement of instrument:** 0 particles/cm³ in 5 minutes

Results (using pulse output):

Particle size (nm)	40	30	20	15	10
Number concentration (cm-3)	1304	1587	1198	1103	921
Counting efficiency η	1.03	1.02	1.01	0.94	0.79
Particle size (nm)	09	08	07	06	05
Number concentration (cm-3)	1032	1008	-	-	-
Counting efficiency η	0.72	0.48	-	-	-
Particle size (nm)	40				
Number concentration (cm-3)	-				
Counting efficiency η	-				

SWIFT CODE: COBADEFF 860











Leibniz Institute for Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Special Information regarding to the Candidate:

Was it necessary to:	yes/no	information
do a second run	no	-
clean the optics	no	-
clean the nozzle	no	-
clean the saturator	no	-
change the wick	no	-
change the laser	no	-
change internal settings	no	-




Fig. 1: Counting efficiency for PALAS CPC ENVI S/N 9532 against aerosol electrometer 3068 S/N 70838596; silver particles between 5 and 40 nm were used for calibration; the calculated Dp50 is 8.03 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB
from display	-	-	-	-	-
Status	P OR	P NO	Laser	LV	flow
from display	-	-	-	-	0.893

Date of issue: October 08, 2019 Reviewed: TROPOS / Kay Weinhold

Page 4 / 4

http://www.tropos.de

IBAN: DE77 8604 0000 0102 1450 00 SWIFT CODE: COBADEFF 860

Commerzbank Leipzig

KTO 102 14 50

