



World Calibration Centre
for Aerosol Physics

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig



Leibniz Institute for
Tropospheric Research

CPC Model: TSI CPC 3750

CPC Serial Number: 3750200601

Customer: TSI Instruments Ltd.

Description: Calibration of a Condensation Particle Counter (CPC, Model 3750)

Date of Calibration: March 18, 2020

Summary of Intercomparison:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 99% efficiency at 40 nm. The Dp50 is at 6.09nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Certificate / Reference: WCCAP

Date of issue: March 20, 2020

Signature:

Reviewed by: **TROPOS**

Name: **Kay Weinhold**

Page 1 / 4



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research

Date of arrival of instrument in calibration lab: *March 16, 2020*
Instrument: *Condensation Particle Counter*
Model and serial number of instrument: *CPC 3750 S/N 375200601*

Result of physical inspection: *no damages*
Result of functional test: *functional test successful, no problems*

Internal parameters of instrument *nominal flow rate 1.0 l/min*

Model and identification number of aerosol electrometer: *TSI Electrometer Model 3068, S/N 70838596*

Electrometer calibration certificate: *September 5, 2018, calibrated at PTB Braunschweig*

Corrections of electrometer, for instance, differing flow rate: *Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.000 l/min*

Software for recording: *LabView 2010; National Instruments; Program „LabCount.vi“*

Date of calibration: *March 18, 2020*
Lab temperature and pressure: *23.0°C, 1008 mbar*
Measured aerosol flow rate of CPC: *0.999 l/min*
Uncertainty in measured flow rate: *3%*
Flowmeter used: *Gilian Gilibrator V; S/N 1711008-S, January, 2018*

Particles and gases used for calibration: *silver particles and nitrogen*
Method of particle generation: *tube furnace generator*
Zero measurement of instrument: *0 particles/cm³ in 5 minutes*

Results (using pulse output and logging via TROPOS Labview software):

Particle size (nm)	40	30	20	15	10	09
Number concentration (cm-3)	1315	1487	991	1197	1644	1608
Counting efficiency η	0.97	0.99	0.99	0.99	0.90	0.85
Particle size (nm)	08	07	06	05	40	
Number concentration (cm-3)	1426	1420	804	357	1071	
Counting efficiency η	0.76	0.64	0.47	0.23	0.98	



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research




Fig. 1: Counting efficiency for CPC 3750 S/N 3750200601 against aerosol electrometer 3068 S/N 70838596; silver particles between 5 and 40 nm were used for calibration; the calculated D_{p50} is 6.09 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB	P VAC
from display	39.0	18	40.0	23.8	101.6	77.3
Status	P OR	P NO	Laser	LV	flow	P INLET
from display	74.5	2.43	41	full	0.999	-0.3

Results (using pulse output and logging via TROPOS Labview software):

without coincidence correction					
Concentration EM in #/cm ³	59180	52207	40447	30357	21658
Number concentration without coincidence correction (cm-3)	46224	41463	33363	25900	19096
Counting efficiency η	0.78	0.79	0.83	0.85	0.88
Concentration EM in #/cm ³	11176	8737	3986	2053	
Number concentration without coincidence correction (cm-3)	10875	8206	3836	2002	
Counting efficiency η	0.92	0.94	0.96	0.98	



World Calibration Centre
for Aerosol Physics



Leibniz Institute for
Tropospheric Research

Results (using USB-C connection and logging via TSI software):

with coincidence correction					
Concentration EM in $\#/cm^3$	59180	52207	40447	30357	21658
Number concentration with coincidence correction (cm^{-3})	60093	52226	40091	30037	21457
Counting efficiency η	1.02	1.00	0.99	0.99	0.99
Concentration EM in $\#/cm^3$	11176	8737	3986	2053	
Number concentration with coincidence correction (cm^{-3})	11177	8777	4015	2076	
Counting efficiency η	1.00	1.00	1.01	1.01	




Fig. 2: Linearity test for TSI CPC 3750 SN 3750200601 against aerosol electrometer 3068 SN 70838596; silver particles with a diameter of 30 nm were used for number concentrations between 2000 and 60000 particles per cm^3 .

Date of issue: March 20, 2020

Reference: TSI electrometer, model 3068, SN 70838596

Reviewed: TROPOS / Kay Weinhold

Page 4 / 4