



Intercomparison of integrating nephelometers Project No.: IN-2020-1-3

Basic informations:

Location of the quality assurance: TROPOS, Lab 121

Date: 27 January - 31 January 2020

Principal Investigator	Home Institution	Participant	Instrument
C. Di Biagio	LISA	C. Di Biagio	1041

1 Intercomparison summary

Status on arrival

No issues due to transportation or other damages.

Zerocheck

The noise level of the instrument is out of the normal range. The average noise (1σ) for the all wavelengths was less equal 0.55 Mm^{-1} for full scattering and 0.37 Mm^{-1} for backscattering. The background level was acceptable with deviations of less equal 0.47 Mm^{-1} for full scattering and 0.41 Mm^{-1} for backscattering.

Spancheck

The span check was unacceptable with deviations of less equal 13.1 %.

Inspection

The measuring cell and the light trap was contaminated with dust. The measuring cell and the light trap was completely cleaned. A recalibrations was not performed.

Comparison to reference nephelometer

Before inspection and recalibration

The results from intercomparisson to reference device were barely acceptable with deviations in the range of -1.8% to 8.6% .

After inspection and recalibration

The results from intercomparisson to reference device were acceptable with deviations in the range of -3.7% to 5.6% .

Recommendations

No recommendations.

Overall assessment

The instrument meets the requirements.

2 Details

Configuration parameters

Zerocheck

Table 1: Noise parameters of nephelometer (SN 1041) measured with filtered air.

Wavelength [nm]	total scattering		backscattering	
	mean [Mm ⁻¹]	std.dev. [Mm ⁻¹]	mean [Mm ⁻¹]	std.dev. [Mm ⁻¹]
450	-0.12	0.55	0.41	0.37
530	0.47	0.26	0.02	0.16
700	0.06	0.31	0.27	0.23

Spancheck

Table 2: Percentage deviation of measured values from nephelometer (SN 1041) to theoretical values for CO₂

Wavelength [nm]	total scattering	backscattering
	deviation [%]	deviation [%]
450	4.5	0
530	7.6	-2.7
700	3.9	-13.1

Comparison to reference nephelometer before inspection and recalibration

Table 3: Comparison of nephelometer (SN 1041) to reference nephelometer Aurora4000 (SN 14-1408) before inspection and recalibration. Testaerosol is ammonium sulfate.

Wavelength [nm]	total scattering slope	total scattering R2	backscattering slope	backscattering R2
450	1.04	1	0.982	0.997
525	1.076	1	1.041	0.998
635	1.032	0.999	1.086	0.997




Figure 1: Correlation of scattering coefficients from nephelometer (SN 1041) and reference nephelometer Aurora4000 (SN 14-1408) before inspection and recalibration. Testaerosol is ammonium sulfate.

Comparison to reference nephelometer after inspection and recalibration

Table 4: Comparison of nephelometer (SN 1041) to reference nephelometer Aurora4000 (SN 14-1408) after inspection and recalibration. Testaerosol is ammonium sulfate.

Wavelength [nm]	total scattering slope	total scattering R2	backscattering slope	backscattering R2
450	1.02	0.999	0.963	0.996
525	1.056	0.999	1.024	0.997
635	1.015	0.999	1.056	0.994




Figure 2: Correlation of scattering coefficients from nephelometer (SN 1041) and reference nephelometer Aurora4000 (SN 14-1408) after inspection and recalibration. Testaerosol is ammonium sulfate.