

Intercomparison of Mobility Particle Size Spectrometers

Project No.: MPSS-2017-2-5

Principal Investigator: Dr. Angela Marinoni

Home Institution: ISAC - CNR
Via Gobetti 101
40129 Bologna
Italy

Participant: -

Candidate: **I-CNR Bologna**
Made by: **TROPOS**
Counter (SN): TSI CPC Model 3772, SN: 70902011
Software: TROPOS

Location of the quality assurance: TROPOS Leipzig, lab 118

Comparison period: June 23, 2017 – June 26, 2017

Last Intercomparison (with Project No.):

Summary of Intercomparison:

Status:


The ISAC-CNR station Bologna is now operating with a TROPOS MPSS with the newest ACTRIS standards. During the night runs from June 23 to June 26, the performance of the system showed a concentration 1% lower than the TROPOS Reference Instrument No.1. The PSL check showed a correct peak at 203.65 nm. The candidate used the calibrated TSI CPC model 3772 and a Ni63 source from TROPOS. The candidate passed the quality standards of ACTRIS and GAW.

Information about the instruments: TROPOS MPSS

Date of check: June 23, 2017

List of Components	TROPOS Reference MPSS No.1	TROPOS Reference MPSS No.	Candidate
Position	Line 1	-	Line 1
Company	TROPOS	-	TROPOS
Software	TROPOS	-	TROPOS V6.68
CPC-MPSS	TSI CPC, Model 3772	-	TSI CPC, Model 3772
CPC-total	TSI CPC, Model 3010	-	-
flow ratio	1.0 : 5.0	-	1.0 : 5.0
source	Kr85	-	Ni63
HV power supply	Positive	-	positive
DMA	Hauke medium	-	Hauke medium
aerosol dryer	✓	-	✓
aerosol RH- sensor	✓	-	✓
aerosol T-sensor	✓	-	✓
sheath RH-sensor	✓	-	✓
sheath T-sensor	✓	-	✓
Sheath dryer	✓	-	✓
pressure sensor	✓	-	✓
info			new system

Laboratory setup:



Status of the instruments:

Date of check (Status): June 23, 2017


<i>CPC status</i>	MPSS		Total CPC	
<i>power/status</i>	LED green	-	-	-
<i>saturator temp</i>	39.0	°C	-	°C
<i>condenser temp</i>	22	°C	-	°C
<i>optics temp</i>	40.1	°C	-	°C
<i>cabinet temp</i>	28.8	°C	-	°C
<i>ambient pressure</i>	99.0	kPa	-	kPa
<i>orifice pressure</i>	72.6	kPa	-	kPa
<i>nozzle pressure</i>	2.8	kPa	-	kPa
<i>laser current</i>	42	mA	-	mA
<i>liquid level</i>	full	-	-	-

Date of system checks:

<i>date</i>	23.06.2017				unit
<i>total CPC flow</i>	-	-			l/min
<i>aerosol flow (DMA)</i>	-	-			l/min
<i>aerosol flow (UDMA)</i>	-	-			l/min
<i>aerosol flow (total)</i>	1.03	-			l/min
<i>Zero MPSS</i>	0	-			#/cm ³
<i>Zero total CPC</i>	-	-			#/cm ³
<i>PSL 203 nm</i>	203.65	-			nm
<i>HV – 0 V</i>	0.2	-			V
<i>HV – 5 V</i>	5.1	-			V
<i>HV – 100 V</i>	100.2	-			V
<i>HV – 1000 V</i>	1000.3	-			V

Special Information regarding the Candidate: I-CNR Bologna is a new TROPOS MPSS

Was it necessary to:	yes/no (date)	old part (ID/SN)	new part (ID/SN)	information
clean the aerosol inlet	no	-	-	-
change aerosol Nafion dryer	no	-	-	-
change sheath Nafion dryer	no	-	-	-
check source	no	-	-	-
change HV power supply	no	-	-	-
clean/change DMA	no	-	-	-
change aerosol RH/T-sensor	no	-	-	-
change sheath RH/T-sensor	no	-	-	-
change pressure sensor	no	-	-	-
change inlet Nafion dryer (500)	no	-	-	-
Change Total filter	no	-	-	-

PSL Scan and calibration: Latex 203 nm +/- 4 nm**Figure 01:** Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on June 15th, 2017.




Figure 02: Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on June 15th, 2017.

Status of the Candidate: Particle Number Size Distribution




Figure 03: Comparison of mean particle number size distribution of TROPOS Reference MPSS No.1 against I-CNR Bologna from June 23, 2017 06:00 PM – June 26, 2017 06:00 AM. Multiple charge correction, internal diffusion losses and CPC efficiency are included.

Status of the Candidate: Time Series




Figure 04: Time series (June 23, 2017 06:00 PM – June 26, 2017 06:00 AM) of the integrated particle number concentration ($N_{10-800nm}$) of the MPSS and total number concentration (N_{total}) of the Reference TSI-CPC Model 3010. The inversion for the candidate was performed using TSI software. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

Status of the Candidate: Correlation




Figure 05: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and TROPOS Reference MPSS No.1 (Night data: June 23, 2017 06:00 PM – June 24, 2017 06:00 AM; June 24, 2017 06:00 PM – June 25, 2017 06:00 AM; June 25, 2017 06:00 PM – June 26, 2017 06:00 AM). Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 06: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and I-CNR Bologna (Night data: June 23, 2017 06:00 PM – June 24, 2017 06:00 AM; June 24, 2017 06:00 PM – June 25, 2017 06:00 AM; June 25, 2017 06:00 PM – June 26, 2017 06:00 AM). Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 07: Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and I-CNR Bologna (Night data: June 23, 2017 06:00 PM – June 24, 2017 06:00 AM; June 24, 2017 06:00 PM – June 25, 2017 06:00 AM; June 25, 2017 06:00 PM – June 26, 2017 06:00 AM). Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 08: Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and I-CNR Bologna (June 23, 2017 06:00 PM – June 26, 2017 06:00 AM). Multiple charge correction, internal diffusion losses and CPC flow corrections are included.