

Intercomparison of Absorption Photometers Project No.: AP-2016-1-1

Basic Information:

Location of the quality assurance: TROPOS, lab 121

Date: 18 July, 2016

| Principal Investigator | Home Institution | Participant | Instrument |
|------------------------|-------------------------|-------------|-------------------------|
| Martin Gysel | Paul Scherrer Institute | - | AE33, SN AE33-S02-00248 |

1. Intercomparison summary

Flow calibration: The flow of the instrument was found to be 4% too low. This deviation is small and could be caused that the instrument calibration point at Jungfraujoch (650 hPa) and Leipzig (1000 hPa) are different. Corrections for the flow deviation and the temperature and pressure (STP correction) were considered in the data evaluation.

Noise. The noise level of the instrument is in the normal range. The average noise (1σ) for all seven wavelengths was 31 ng/m^3 for 1 minute averaging time. The maximum noise was 38.7 ng/m^3 at 370 nm.

Inspection: Measurement cell was clean. The spots edges are sharp.

Comparison to reference MAAP: BC concentrations at wavelength 660 nm are 2% higher than BC concentrations from a reference MAAP (SN 504).

Comparison to reference Aethalometer AE33 (SN S02-00163): The AE33 (SN 248) measures at all wavelengths lower concentrations than the reference Aethalometer of type AE33 (SN 163). The slopes in the correlation plot are between 0.921 (950 nm) and 0.947 (370 nm).

Comparison to reference absorption: Absorption coefficients derived from Aethalometer using a conversion factors of $C_0 = 1.63$ are higher by factors 1.267 (470nm), 1.187 (520 nm), and 1.096 (660) nm than absorption coefficients from the reference absorption setup. Note, there is no agreement on a common conversion factor for Aethalometers of type AE33. Furthermore, the conversion factor can vary depending on aerosol type.

Recommendations: None.

Overall assessment: The instrument meets the requirements.

2. Details

| Configuration parameters |
|--|
| Instrument serial number: S02-00248 |
| BC Unit: 0 (ng) |
| Sigma values: 18.47, 14.54, 13.13, 11.58, 10.35, 7.77, 7.19 |
| Volumetric reference: 'Standard' with $P_0=650$ hPa and $T_0=25^\circ\text{C}$ |

Data processing

Equivalent black carbon concentrations reported by instruments were corrected for flow, spot size deviations and adjusted to standard temperature and pressure conditions ($T=0^\circ\text{C}$, $P=1013.25$ hPa) by

$$[BC] = [BC_{instr}] \times F_{flow} \times F_{spot} \times F_{STP}$$

For details read Appendix A.

Conversions between eBC concentrations and absorption coefficients are done by

$$b_{abs}[1/Mm] = eBc[\mu g/cm] \times \sigma_{air}/C_0 ,$$

with $C_0=1.63$ mass absorption cross sections σ_{air} given in the table below. For individual instruments the Sigma-values can be found in the setup file.

| Conversion factors (σ_{air}) for eBC concentrations to absorption coefficients | | | | | | | |
|---|-------|-------|-------|-------|-------|------|------|
| Wavelength [nm] | 370 | 450 | 530 | 590 | 660 | 880 | 950 |
| σ_{air} [m^2/g] | 18.47 | 14.54 | 13.13 | 11.58 | 10.35 | 7.77 | 7.19 |

Flow check

Correction factors F_{flow} and F_{STP} for correcting eBC concentrations. F_{flow} corrects for inlet flow errors. F_{STP} is used to adjust concentrations to STP conditions (0°C, 1013.25 hPa).

| Date | System Flow | | | Reference flow | | | Flow correction factor <small>Fehler! Textmarke nicht definiert.</small> | STP correction factor Fehler! Textmarke nicht definiert. |
|---------|----------------------|----------------------|-----------------------|--|---------------------|--------------|---|--|
| | | | | Reference flow meter: Gilibrator ‘TROPOS-T’ | | | | |
| | Mass flow | Volume reference | | Volume flow | Ambient T and P | | | |
| | Q_{AE33} [slpm] | $T_{0,AE33}$ [°C] | $P_{0,AE33}$ [hPa] | Q [lpm] | T [°C] | P [hPa] | | |
| 21. Sep | 5 | 25 | 650 | 3.37 | 22 | 997 | 0.957 | 1.701 |

Spot size check

Correction factor for spot sizes F_{spot} .

| Date | Nominal spot size [cm ²] | Measured spot size [mm ²] | F_{spot} |
|-----------|--------------------------------------|---|------------|
| 2016-03-8 | 0.785 | Well defined spot, spot size not measured | 1.0 |

Instrumental Noise

Noise in units of eBC concentration measured with filtered air.

| Date | Avg. time | Wave-length [nm] | Num data points | Median [ng] | 10 th percentile [ng/m ³] | 90 th percentile [ng/m ³] | Mean [ng/m ³] | Standard deviation [ng/m ³] | Error of the mean [ng/m ³] |
|----------|-----------|------------------|-----------------|-------------|--|--|---------------------------|---|--|
| March 10 | 5 min | 370 | 393 | 2 | -9 | 13 | 3.9 | 38.7 | 2 |
| | | 450 | 393 | -1 | -16 | 13 | 0.3 | 32.6 | 1.6 |
| | | 520 | 393 | -2 | -18 | 15 | -0.4 | 29 | 1.5 |
| | | 590 | 393 | -2 | -21 | 19 | -0.5 | 29.3 | 1.5 |
| | | 660 | 393 | -2 | -18.7 | 15 | -0.8 | 28.5 | 1.4 |
| | | 880 | 393 | -2 | -21 | 19 | -0.1 | 29.1 | 1.5 |
| | | 950 | 393 | -2 | -21.7 | 17.7 | -0.7 | 30.9 | 1.6 |

Comparison of AE33 and MAAP

Correlation of eBC from AE33 (SN 248) at 660 nm from the reference MAAP (SN 504) at 637 nm

| Wavelength [nm] | AE33: 660 nm MAAP: 637 nm |
|-----------------|------------------------------|
| Slope | 1.021±0.002 |
| R ² | 0.932 |




Figure 1: Comparison of eBC concentration for AE33 SN-248 (660 nm) and MAAP SN-504 (637 nm).

| Comparison of AE33 (SN 248) to the reference AE33 (SN163) | | | | | | | |
|---|-------------|-------------|-------------|-------------|-------------|-------------|-------------|
| Wavelength [nm] | 370 | 470 | 520 | 590 | 660 | 880 | 950 |
| Slope | 0.997±0.002 | 0.992±0.002 | 0.986±0.002 | 0.977±0.015 | 0.943±0.002 | 0.954±0.002 | 0.937±0.002 |
| R ² | 0.947 | 0.946 | 0.946 | 0.946 | 0.933 | 0.924 | 0.921 |




Figure 2: Comparisons of AE33 (SN-248) to the reference instrument AE33 (SN163) at 880 nm.

| Comparison of absorption coefficients from AE33 (SN 408) and the Multi-wavelength reference absorption (Extinction minus Scattering). | | | |
|---|-------------|-------------|-------------|
| Wavelength [nm] | 470 | 520 | 660 |
| Slope | 1.267±0.013 | 1.187±0.010 | 1.096±0.018 |
| R ² | 0.464 | 0.464 | 0.162 |




Figure 3: Comparison for 470 nm. Blue: reference absorption, Red: AE33 (SN-248).




Figure 4: Comparison for 520 nm. Blue: reference absorption, Red: AE33 (SN-248).




Figure 5: Comparison for 660 nm. Blue: reference absorption, Red: AE33 (SN-248).

Appendix: Instrument corrections

Necessary corrections to all instruments are flow and spot size correction and conversion of concentrations and absorption coefficients to STP conditions. BC concentrations from individual instruments $[BC_{instr}]$ were by corrected by:

$$[BC] = [BC_{instr}] \times F_{flow} \times F_{spot} \times F_{STP} \times 1/mean_ratio$$

- a) The Flow correction factor for compensating calibration errors of the instrument flow meter and is defined by:

$$F_{flow} = \frac{Q_{instr} [slpm]}{Q_{ref} [lpm]} \times \frac{T_{ref} [K]}{T_{0,instr} [K]} \times \frac{P_{0,instr} [hPa]}{P_{ref} [hPa]}$$

where $Q_{instr.}$ and Q_{ref} are the flows measured with the instrument and determined with a reference volume flow meter, respectively. The flow of the volume flow meter is converted using the temperature T_{ref} and pressure P_{ref} , which are typically the ambient or room temperature or pressure near the reference flow meter. Also the standard temperature $T_{0,instr}$ and standard pressure $P_{0,instr}$ of the instrument have to be considered.

- b) The adjustment of instrument flow to standard temperature and pressure (STP) is done by

$$F_{STP} = \frac{T_{0,instr.} + 273}{T_0 + 273} \times \frac{P_0}{P_{0,instr.}}$$

- c) whereas $T_{0,instr}$ and $P_{0,instr.}$ are the standard temperature and pressure of individual instrument. For ACTRIS workshops STP is defined to be $T_0=0^\circ\text{C}$ and $P_0=1013.25$ hPa.

- d) The spot size correction factor F_{spot} compensates for systematic deviations of sample spot sizes and is defined by

$$F_{spot} = \frac{A_{meas}}{A_{instr}}$$

where $A_{instr.}$ and A_{meas} are the instrument nominal and the measured spot area, respectively.

- e) The mean ration is a calibration parameter and can be found in the setup file of AE31 instruments. This factory calibration is undone for ACTRIS intercomparisons. If the mean ration deviates from unity, special care must be taken, since this calibration factor is always included in data from Aethalometers and can not be switched off.

This issue must be considered when discussion deviations to reference instruments.